



**KAVALIER**


# DECLARATION OF COMPLIANCE

Issuer's name/producer: **KAVALIERGLASS, a.s.**  
 Issuer's address/Producer: **Křížová 1018/6, Prague 5**  
**office: Sklářská 359, 285 06 Sázava, Czech Republic**

Object of the declaration: **FLASKS VOLUMETRIC**  
**with SJ and plastic stopper, class A**

<u>Product IDN &amp; Description</u>	<u>Catalog No.</u>	<u>Capacity/ ml</u>
1503 APN	1632431622018	5
	1632431622019	10
	1632431622023	25
	1632431622025	50
	1632431622030	100
	1632431622037	200
	1632431622038	250
	1632431622043	500
	1632431622044	1000
	1632431622046	2000

## Glass and Plastic accessories



<b>Material specification:</b>		
<b>Flask volumetric</b>	clear	Borosilicate glass SIMAX®
<b>Print on the body of the flask</b>	white	in fired-on, chemically resistant ceramic enamel
<b>Stopper</b>	white blue target	PE LITEN MB 68 natur PE BRALEN VA 2060
<b>Purpose of use</b>	laboratory volumetric glassware exact measuring of the liquid capacity, preparation and storage of standard solutions	

**The object of the certificate described above is in conformity with the requirements of the following standards and regulations:**

**Glass characteristics:**

- ISO 3585 Borosilicate glass 3.3 – Properties
  - Chemical durability (art. 4.1, 4.2, 4.3, 4.4)
  - Physical properties (art. 5.1, 5.2, 5.3, 5.4, 5.5, 5.6)
- Glass containers for pharmaceutical use
  - Eur. Ph 10<sup>th</sup> -3.2.1 Glass Type I.
- ISO 1042 Laboratory glassware – One-mark volumetric flasks
  - Maximum permissible errors in capacity fulfill the values specified in Table 1/ISO 1042.

Table 1 – ISO 1042

Class A, Certified Volumetric Flasks, Transparent		
Nominal capacity [ml]	Accuracy limits [±ml]	Stopper size NS
5	0,025	7/16
10	0,025	7/16
25	0,040	10/19
50	0,060	12/21
100	0,100	12/21
200	0,150	14/23
250	0,150	14/23
500	0,250	19/26
1000	0,400	24/29
2000	0,600	29/32

**No heavy metals (lead, cadmium, mercury and hexavalent chromium):**

- **Regulation (EC) No. 987/2008 of 8 October 2008 amending Regulation (EC) No. 1907/2006 – REACH as regards Annexes IV and V – glass was exempted from the obligation to register.**

**Chemical characteristics (acc. to Regulation No 1907/2006/EC):**

Composition:	<b>CAS No.</b>	<b>EINECS No.</b>	<b>Component:</b>	<b>Concentration /Percent:</b>
	65997-17-3	266-046-0	Glass, oxide, chemicals	100%

**Characteristics of Borosilicate glass SIMAX®:**

- **Acid resistance** Class I. (to ISO 1776)
- **Hydrolytic resistance** Class I. (HGB1 to ISO 719; HGA1 to ISO 720)
- **Alkali resistance** Class II. (to ISO 695)
- **Coefficient of mean linear thermal expansion  $\alpha$ :  $3,3 \times 10^{-6} \text{ K}^{-1}$  (to ISO 7991; 20/300 °C)**

Additional information:

The producer confirms hereby that the characteristics, measures and accuracy of the products listed above are in full conformity with the provisions of the standard.

The producer also declares that the products are safe when used in usual and proper way.

The producer has installed the Quality Assurance System according to ISO 9001 and thus guarantees that all products delivered to the market are in full conformity with the technical documentation and with all fundamental requirements to such products.  
Certificate No. 04 100 940602 issued by TÜV CERT, Certification Body at TÜV NORD CERT GmbH.

The certificate is issued for the customer: -

Sázava, 29. 01. 2021  
Place and date of issue

Ing. Kristýna Machová  
Project Quality Engineer

**KAVALIERGLASS, a.s.**  
Křížová 1018/6, 150 00 Praha 5  
office: Sklářská 359, 285 06 Sázava  
Czech Republic  
IČ: 474 68/815  
-61-