



**KAVALIER**

# CERTIFICATE OF CONFORMITY

Issuer's name/ producer:

**KAVALIERGLASS, a.s.**

Issuer's address/Producer:

**Křížová 1018/6, Prague 5**

**office: Sklářská 359, 285 06 Sázava, Czech Republic**

Object of the declaration:

**GLASS TUBES; RODS; CAPILLARIES**

## Scheme of the glass item



Material specification:			Technical data/ Declaration of compliance
tube, rod, capillary	clear	Borosilicate glass SIMAX®	Page 2/ Page 3-4
Purpose of use			Application in technical, pharmaceutical, laboratory or food industry

# TECHNICAL REQUIREMENTS ACCORDING TO PURPOSE OF USE

---

## Characteristics of Borosilicate glass SIMAX®

- **Acid resistance** Class I. ISO 1776
- **Hydrolytic resistance** Class I. HGB1 to ISO 719;  
HGA1 to ISO 720
- **Alkali resistance** Class II. ISO 695
- **Coefficient of mean linear thermal expansion  $\alpha$ :  $3,3 \times 10^{-6} \text{ K}^{-1}$**  ISO 7991; (20/300 °C)
- **Pharmaceutical use**

	<i>European Pharmacopoeia (EP)</i>	<i>US Pharmacopoeia (USP)</i>	<i>Japanese Pharmacopoeia (JP)</i>
<b>Glass</b>	Eur. Ph.10 <sup>th</sup> – 3.2.1	USP <660>	JP16

### Supporting data:

TEST / European Pharmacopoeia 10, Art. 3.2.1	UNIT	LIMIT	RESULT
Hydrolytic resistance - inner surfaces, test A	ml 0,01 mol/l HCl/100ml of leachate	max 0,40	0,04
Hydrolytic resistance - glass grains, test B	mol 0,02/l HCl/g	max 0,1	0,038
Arsenic content	$\mu\text{g As/g}$	max 0,1	< 0,001

- **Chemical characteristics (acc. to Regulation No 1907/2006/EC):**

Composition:	<b>CAS No.</b>	<b>EINECS No.</b>	<b>Component:</b>	<b>Concentration /Percent:</b>
	65997-17-3	266-046-0	Glass, oxide, chemicals	100%

**Chemical stability:** Stable

### Chemical characteristics of borosilicate glass (approximate values)

Component	Content (percentage by weight)
SiO <sub>2</sub>	80,3%
B <sub>2</sub> O <sub>3</sub>	13,0%
Al <sub>2</sub> O <sub>3</sub>	2,4%
Na <sub>2</sub> O + K <sub>2</sub> O	4,3%



**KAVALIER**

# DECLARATION OF COMPLIANCE

Issuer's name: **KAVALIERGLASS, a.s.**  
Issuer's address: **Křížová 1018/6, Prague 5**  
**office: Sklářská 359, 285 06 Sázava, Czech Republic**

Object of the declaration: **GLASS TUBES; RODS; CAPILLARIES**

Material: **Borosilicate glass SIMAX®, glass with high thermal and chemical resistance**  
Country of origin: **Czech Republic**  
Purpose of use: **Application in technical, pharmaceutical, laboratory or food industry**

**The object of the certificate described above is in conformity with the requirements of the following standards and regulations:**

## **GLASS CHARACTERISTICS:**

- ISO 3585 Borosilicate glass 3.3 – Properties
  - Chemical durability (art. 4.1, 4.2, 4.3, 4.4)
  - Physical properties (art. 5.1, 5.2, 5.3, 5.4, 5.5, 5.6)
- Glass containers for pharmaceutical use
  - Eur. Ph 10<sup>th</sup> Edition - 3.2.1 Glass Type I.

## **FOOD CONTACT:**

- Commission Regulation (EU) No. 2023/2006

Good manufacturing practice for materials and articles intended to come into contact with food

- Regulation EC No 1935/2004 of 27 October 2004

Directive on materials and articles intended to come into contact with food and repealing Directives 80/590/EEC and 89/109/EEC

- Regulation of Czech Health Ministry Decree No. 38/2001 Coll.

Directive on articles intended to come into contact with foodstuffs

- Directive 84/500EEC of 15 October 1984

Directive on the approximation of the laws of the Member States relating to ceramic articles intended to come into contact with foodstuffs.

- ISO 7086-1:2000 Glass hollowware in contact with food

Release of lead and cadmium – Part 1: Test method

- ISO 7086-2:2000 Glass hollowware in contact with food

Release of lead and cadmium – Part 2: Permissible limits

**No heavy metals (lead, cadmium, mercury and hexavalent chromium):**

- Regulation (EC) No. 1907/2006 of the European Parliament and of the Council of 18<sup>th</sup> December – [EU REACH Regulation](#)

Products do not release Substances of Very High Concern (SVHCs) above their respective threshold values listed in SVHC Candidate List, published by European Chemicals Agency (ECHA) as a part of REACH Regulation with latest publication date 22.01.2021.

- RoHS

Directive 2011/65/EC (RoHS II), amended by 2015/863/EC (RoHS III), on the restriction of the use of certain hazardous substances in electrical and electronic equipment, Annex II - extension of limitation regarding 4 additional substances.

- CMR/ ED substances

We hereby certify that the borosilicate glass SIMAX® does not contain any of the CMR/ ED substances.

Additional information:

---

**The producer confirms hereby that the characteristics, measures and accuracy of the products listed above are in full conformity with the provisions of the standard.**

**The producer also declares that the products are safe when used in usual and proper way.**

The producer has installed the Quality Assurance System according to ISO 9001 and thus guarantees that all products delivered to the market are in full conformity with the technical documentation and with all fundamental requirements to such products.  
Certificate No. 04 100 940602 issued by TÜV CERT, Certification Body at TÜV NORD CERT GmbH.

---

The certificate is issued for the customer: -

Sázava, 26. 05. 2021  
*Place and date of issue*

Ing. Kristýna Machová  
Project Quality Engineer

**KAVALIERGLASS, a.s.**  
Křížová 1018/6, 150 00 Praha 5  
office: Sklářská 359, 285 06 Sázava  
Czech Republic  
IC: 474 68/815  
-61-

